



**National Accreditation Board for
Testing and Calibration Laboratories**

(A Constituent Board of Quality Council of India)



CERTIFICATE OF ACCREDITATION

ELECTRONICS TEST & DEVELOPMENT CENTRE

has been assessed and accredited in accordance with the standard

ISO/IEC 17025:2005

"General Requirements for the Competence of Testing & Calibration Laboratories"

for its facilities at

30, GMC-NIO Road, P.O. Goa University, Goa

in the field of

CALIBRATION

Certificate Number CC-2049

Issue Date 29/11/2018

Valid Until 28/11/2020

This certificate remains valid for the Scope of Accreditation as specified in the annexure subject to continued satisfactory compliance to the above standard & the relevant requirements of NABL.

(To see the scope of accreditation of this laboratory, you may also visit NABL website www.nabl-india.org)

Signed for and on behalf of NABL



89076970200020000669

Anil Relia

Anil Relia
Chief Executive Officer



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Electronics Test & Development Centre, 30, GMC-NIO Road,
P.O. Goa University, Goa

Accreditation Standard ISO/IEC 17025: 2005


Certificate Number CC-2049

Page 1 of 10

Validity 29.11.2018 to 28.11.2020

Last Amended on -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability ($\pm$)	Remarks
ELECTRO-TECHNICAL CALIBRATION				
I.	SOURCE			
1.	DC Voltage #	1 mV to 100 mV 100 mV to 10 V 10 V to 1000 V	0.13 % to 0.004 % 0.004 % to 0.002 % 0.002 %	Using Fluke 5520A by Direct Method
2.	DC Current #	100 μ A to 1 A 1 A to 20 A 20 A to 1000 A	0.04 % to 0.027 % 0.027 % to 0.12 % 0.4 % to 0.5 %	Using Fluke 5520A by Direct Method Using Fluke 5520A, 50 Turn Current Coil by Direct Method
3.	Resistance #	1 m Ω to 10 m Ω 0.01 Ω to 1 Ω 1 Ω to 100 Ω 100 Ω to 1 M Ω 1 M Ω to 10 M Ω 10 M Ω to 1 G Ω	0.70 % to 0.088 % 0.088 % to 0.01 % 0.01 % to 0.005 % 0.005 % to 0.01 % 0.012 % 0.01 % to 0.57 %	Using Vaiseshika 9409B, Fluke 5520A, ESI SR1 Series, ZMSR Series by Direct Method
4.	AC Voltage #	50 Hz to 8 kHz 10 mV to 100 mV 100 mV to 10 V 10 V to 1000 V	0.10 % to 0.03 % 0.03 % 0.03 % to 0.04 %	Using Fluke 5520A by Direct Method
5.	AC Current #	50 Hz to 5 kHz 100 μ A to 100 mA 100 mA to 3 A 3 A to 20 A 50 Hz 20 A to 1000 A	0.2 % to 0.08 % 0.08 % to 0.1 % 0.1 % to 0.17 % 0.39 % to 0.71 %	Using Fluke 5520A by Direct Method Using Fluke 5520A, 50 Turn Current Coil by Direct Method


Pankaj Varshney
Convenor


Avijit Das
Program Manager



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Electronics Test & Development Centre, 30, GMC-NIO Road,
P.O. Goa University, Goa

Accreditation Standard ISO/IEC 17025: 2005


Certificate Number CC-2049

Page 2 of 10

Validity 29.11.2018 to 28.11.2020

Last Amended on -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability ($\pm$)	Remarks
6.	Capacitance [#]	1 kHz 100 pF to 1000 pF 1000 pF to 1 μ F 1 μ F to 10 mF	0.16 % to 0.05 % 0.05 % 0.05 % to 1.8 %	Using GR STD Capacitors, Fluke 5520A, Fluke 9100 by Direct Method
7.	Inductance [#]	1 kHz 100 μ H to 1 mH 1 mH to 10 H	0.4 % to 0.11 % 0.11 % to 0.12 %	Using GR STD Inductors (Discrete Values in Decades) by Direct Method
8.	Frequency [#]	1 Hz to 10 kHz 10 kHz to 100 MHz 100 MHz to 3 GHz	0.0021 % to 0.00002 % 0.00002 % to 0.00001 % 0.00001 %	Using Fluke 5520A, R&S SMB 100A by Direct Method
9.	Oscilloscope Level, Time Marker, Bandwidth [#]	1 mVpp to 100 Vpp 5 s to 2 ns 10 kHz to 600 MHz	0.4 % to 0.14 % 0.05 % to 0.02 % 4 %	Using Fluke 5520A with Scope Option by Direct Method
10.	Temperature Simulation [#]			Using Eurotron Microcal 200+ in Source Mode by Simulation Method
	RTD Pt-100	(-) 200 °C to 800 °C	0.15 °C	
	J Type Thermocouple	(-) 190 °C to 1200 °C	0.25 °C	
	K Type Thermocouple	(-) 100 °C to 1260 °C	0.28 °C	
	T Type Thermocouple	(-) 100 °C to 400 °C	0.16 °C	
	R Type Thermocouple	150 °C to 1750 °C	0.44 °C	
	S Type Thermocouple	170 °C to 1750 °C	0.44 °C	
	B Type Thermocouple	920 °C to 1800 °C	0.54 °C	
	E Type Thermocouple	(-) 190 °C to 1000 °C	0.23 °C	
	N Type Thermocouple	100 °C to 1300 °C	0.26 °C	


Pankaj Varshney
Convenor


Avijit Das
Program Manager



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Electronics Test & Development Centre, 30, GMC-NIO Road,
P.O. Goa University, Goa

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2049

Page 3 of 10

Validity 29.11.2018 to 28.11.2020

Last Amended on -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability ($\pm$)	Remarks
II.	MEASURE			
1.	DC Voltage #	1 mV to 100 mV 100 mV to 10 V 10 V to 1000 V	0.035 % to 0.001 % 0.001 % 0.0012 %	Using HP 3458A DMM by Direct Method
2.	DC Current #	100 μ A to 100 mA 100 mA to 1 A 1 A to 20 A	0.004 % 0.004 % to 0.014 % 0.014 % to 1.14 %	Using HP 3458A DMM by Direct Method Using Shunt UDL4-Z2 with DMM by V/R Method
3.	Resistance #	1 m Ω to 10 m Ω 0.01 Ω to 1 Ω 1 Ω to 100 Ω 100 Ω to 10 M Ω 10 M Ω to 100 M Ω 100 M Ω to 1 G Ω	5.74 % to 0.57 % 0.57 % to 0.008 % 0.008 % to 0.002 % 0.002 % to 0.009 % 0.009 % to 0.05 % 0.05 % to 0.572 %	Using HP 3458A DMM by Direct Method
4.	AC Voltage #	50 Hz to 1 kHz 10 mV to 100 mV 100 mV to 1 V 1 V to 700 V 50 Hz 700 V to 1000 V	0.036 % to 0.01 % 0.01 % 0.01 % to 0.07 % 0.07 % to 0.13 %	Using HP 3458A DMM by Direct Method Using Voltech PM 3000A by Direct Method
5.	AC Current #	50 Hz to 1 kHz 100 μ A to 100 mA 100 mA to 1 A 1 A to 20 A	0.1 % to 0.09 % 0.09 % to 0.11 % 0.11 % to 0.6 %	Using HP 3458A DMM by Direct Method Using Shunt UDL4-Z2 with DMM by V/R Method

Pankaj Varshney
Convenor

Avijit Das
Program Manager



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Electronics Test & Development Centre, 30, GMC-NIO Road,
P.O. Goa University, Goa

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2049

Page 4 of 10

Validity 29.11.2018 to 28.11.2020

Last Amended on -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability ($\pm$)	Remarks
6.	DC High Voltage #	1 kV to 30 kV	2.5 % to 1.5 %	Using Fluke 80K40 / Fluke 87V by Direct Method
7.	AC High Voltage #	50 Hz 1 kV to 15 kV	5.92 %	Using Fluke 80K40 / Fluke 87V by Direct Method
8.	Capacitance #	1 kHz 100 pF to 1 μ F	0.15 % to 0.07 %	Using HP 4274 A LCR Meter by Direct Method
9.	Inductance #	1 kHz 100 μ H to 10 H	0.31 % to 0.155 %	Using HP 4274 A LCR Meter by Direct Method
10.	Frequency #	10 Hz to 10 kHz 10 kHz to 3.0 GHz	0.00012 % to 0.00006 % 0.00006 % to 0.000002 %	Using CNT 90 Pendulum Counter Timer by Direct Method
11.	Temperature Simulation #			
	RTD Pt-100	(-) 200 °C to 800 °C	0.15 °C	Using Eurotron Microcal 200+ in Measure Mode by Simulation Method
	J Type Thermocouple	(-) 190 °C to 1200 °C	0.26 °C	
	K Type Thermocouple	(-) 100 °C to 1350 °C	0.37 °C	
	T Type Thermocouple	(-) 190 °C to 400 °C	0.17 °C	
	R Type Thermocouple	145 °C to 1750 °C	0.44 °C	
	S Type Thermocouple	165 °C to 1750 °C	0.44 °C	
	B Type Thermocouple	915 °C to 1800 °C	0.55 °C	
	E Type Thermocouple	(-) 190 °C to 1000 °C	0.41 °C	
	N Type Thermocouple	1 °C to 1300 °C	0.27 °C	

Pankaj Varshney
Convenor

Avijit Das
Program Manager



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Electronics Test & Development Centre, 30, GMC-NIO Road,
P.O. Goa University, Goa

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2049

Page 5 of 10

Validity 29.11.2018 to 28.11.2020

Last Amended on -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability ($\pm$)	Remarks
12.	Time Interval #	1 s to 3600 s	0.035 s to 0.04 s	Using CNT 90 Pendulum Counter Timer by Comparison Method

Pankaj Varshney
Convenor

Avijit Das
Program Manager



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Electronics Test & Development Centre, 30, GMC-NIO Road,
P.O. Goa University, Goa

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2049

Page 6 of 10

Validity 29.11.2018 to 28.11.2020

Last Amended on -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability ($\pm$)	Remarks
MECHANICAL CALIBRATION				
I.	DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)			
1.	External Micrometer Analog/Digital ^{\$} L.C.: 0.01 mm	0 to 25 mm	6.5 μ m	Using Gauge Blocks Grade "O"
2.	Vernier Calipers Analog/Dial/Digital ^{\$} L.C.: 0.01 mm	0 to 300 mm	9.0 μ m	Using Gauge Blocks Grade "O"
3.	Dial Gauge: Plunger Type Analog/ Digital ^{\$} L.C.: 0.01 mm	0 to 20 mm	7.0 μ m	Using Gauge Blocks Grade "O" & Comparator Stand
4.	Vernier Height Gauge Analog/Dial/Digital ^{\$} L.C.: 0.02 mm	0 to 600 mm	16.0 μ m	Using Slip Gauges, Long Slip Gauges & Surface Plate
II.	WEIGHTS			
1.	Mass/Weights ^{\$}	1 mg 2 mg 5 mg 10 mg 20 mg 50 mg 100 mg 200 mg 500 mg 1 g 2 g	0.05 mg 0.05 mg 0.05 mg 0.05 mg 0.05 mg 0.05 mg 0.05 mg 0.05 mg 0.06 mg 0.06 mg 0.06 mg	Using E2 Class Standard Weights & Digital Semi Micro Balance L.C. 0.01/0.1 mg for M1 Class Weights and Coarser

Pankaj Varshney
Convenor

Avijit Das
Program Manager



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Electronics Test & Development Centre, 30, GMC-NIO Road,
P.O. Goa University, Goa

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2049

Page 7 of 10

Validity 29.11.2018 to 28.11.2020

Last Amended on -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability ($\pm$)	Remarks
		5 g 10 g 20 g 50 g 100 g 200 g	0.06 mg 0.06 mg 0.08 mg 0.15 mg 0.15 mg 0.15 mg	
III.	PRESSURE INDICATING DEVICES			
1.	Pneumatic Pressure Pressure Gauges, Pressure Indicators, Pressure Calibrators ^{\$}	(-) 0.85 bar to 0 bar 0 to 25 bar	0.0065 bar 0.013 bar	Using Pneumatic Digital Pressure Calibrator by Comparison Method as per DKD-R-6
2.	Hydraulic Pressure Pressure Gauges, Pressure Indicators, Pressure Calibrators ^{\$}	0 to 400 bar	0.183 bar	Using Pressurement Dead Weight Tester by Comparison Method as per DKD-R-6
3.	Pneumatic Pressure Pressure Gauges, Pressure Indicators, Pressure Calibrators [*]	(-) 0.85 bar to 0 bar 0 to 25 bar	0.0068 bar 0.139 bar	Using Pneumatic Digital Pressure Calibrator by Comparison Method as per DKD-R-6
4.	Hydraulic Pressure Pressure Gauges, Pressure Indicators, Pressure Calibrators [*]	0 to 400 bar	0.38 bar	Using Digital Hydraulic Pressure Calibrator by Comparison Method as per DKD-R-6

Pankaj Varshney
Convenor

Avijit Das
Program Manager



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Electronics Test & Development Centre, 30, GMC-NIO Road,
P.O. Goa University, Goa

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2049

Page 8 of 10

Validity 29.11.2018 to 28.11.2020

Last Amended on -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability ($\pm$)	Remarks
THERMAL CALIBRATION				
I.	TEMPERATURE			
1.	Glass Thermometers ^s	35 °C to 180 °C	0.32 °C	Using Oil Bath, SSPRT & Temperature Calibrator by Comparison Method
2.	Temperature gauges, RTD / Thermocouple with or without Temp. Indicator/Recorder/ Data Logger ^s	35 °C to 200 °C	0.14 °C	Using Oil Bath, Dry Block, SSPRT & Temperature Calibrator by Comparison Method
		200 °C to 600 °C	0.5 °C	Using Dry Block, SSPRT & Temperature Calibrator by Comparison Method
3.	Thermocouple with or without Temp. Indicator/ Recorder/ Data Logger ^s	600 °C to 1000 °C	1.8 °C	Using "S" Type Thermocouple, Temperature Calibrator, Dry Block Furnace by Comparison Method
4.	Temperature Indicator with Sensor of Baths, Dry Blocks [#]	(-) 35 °C to 50 °C 50 °C to 200 °C 200 °C to 600 °C 600 °C to 1000 °C	0.14 °C 0.24 °C 0.5 °C 1.9 °C	Using SSPRT/ "S" Type Thermocouple, Temperature Calibrator Single Position Calibration (At Measuring Location in DUC)

Pankaj Varshney
Convenor

Avijit Das
Program Manager



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Electronics Test & Development Centre, 30, GMC-NIO Road,
P.O. Goa University, Goa

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2049

Page 9 of 10

Validity 29.11.2018 to 28.11.2020

Last Amended on -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability ($\pm$)	Remarks
5.	Calibration of Environmental Chambers, Air Ovens, Deep Freezers etc.*	(-) 35 °C to 250 °C	1.2 °C	Using PT-100 Sensors (Minimum Nine) with Data Logger at Multi Position Calibration
6.	Temperature Indicator with Sensor of Ovens, Chambers, Deep Freezers*	(-) 35 °C to 250 °C	1.2 °C	Using SSPRT & Temperature Calibrator Single Position Calibration (At Measuring Location in DUC)
7.	Temperature Indicator with Sensor of Ovens, Furnace*	250 °C to 1000 °C	1.9 °C	Using S Type Thermocouple & Temperature Calibrator Single Position Calibration (At Measuring Location in DUC)
II.	SPECIFIC HEAT AND HUMIDITY			
1.	Humidity and Temperature Indicator with Inbuilt or External Sensor, Thermo Hygrometer*	35% RH to 95% RH @25 °C to 35 °C 25 °C to 55 °C 35% RH to 95% RH	2.4 % RH 0.38 °C	Using Digital Thermo Hygrometer, Temperature / Humidity Chamber by Comparison Method

Pankaj Varshney
Convenor

Avijit Das
Program Manager



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Electronics Test & Development Centre, 30, GMC-NIO Road,
P.O. Goa University, Goa

Accreditation Standard ISO/IEC 17025: 2005

Certificate Number CC-2049 **Page** 10 of 10

Validity 29.11.2018 to 28.11.2020 **Last Amended on** -

Sl.	Quantity Measured / Instrument	Range/Frequency	*Calibration Measurement Capability ($\pm$)	Remarks
2.	Humidity Indicator with Sensor of Humidity Chambers, Test Cabinets Etc.*	30% RH to 95% RH @ 25 °C to 55 °C	2.3 % RH	Using Digital Thermo Hygrometer with Probe Single Position Calibration (At Measuring Location in DUC)

* Measurement Capability is expressed as an uncertainty ($\pm$) at a confidence probability of 95%

§ Only in Permanent Laboratory

* Only for Site Calibration

The laboratory is also capable for site calibration however, the uncertainty at site depends on the prevailing actual environmental conditions and master equipment used.

Pankaj Varshney
Convenor

Avijit Das
Program Manager